

AMENDMENT NO. 1

Subject:- Tender No. BPPI/SURGICAL-047/2017 Dated 02.01.2017 for supply of Surgical Consumables to Bureau of Pharma Public Sector Undertakings of India(BPPI) for the year 2017-19

Reference:- Pre-Bid meeting held on 11.01.2017 at 11:00AM in the premises of BPPI

In view of discussion held in pre-bid meeting on 11.01.2017, the following amendment in subject tender document is hereby authorized: -

I. 2.ELIGIBILITY CRITERIA

(a) Add the following after clause 2 (a)(ii) :-

The distributor should have valid sale license.

(b) Add the following after clause 2 (a)(iv) :-

The Marketer should have valid sale license.

II. 4. TECHNICAL BID – COVER “A”

(a) Add the following after clause 4.1 (a) (ii):-

In case, bidder is a marketer of a manufacturer, scanned copy of valid sale license is required to be uploaded.

(b) Add the following after clause 4.1 (a) (ii):-

In case, bidder is a distributor of a manufacturer, scanned copy of valid sale license is required to be uploaded.

(c) Add the following after clause 4.1(f) (i) & (ii):-

In case, bidder is a distributor of a manufacturer or more than one manufacturer, the certificate from the C.A.(Chartered Accountant) or ICWA for Production & financial capacity to manufacture as per format (ANNEXURE V) and Average Annual Turnover certificate from Chartered Accountant of manufacturer (ANNEXURE-VI) are required of their manufacturer/manufacturers.

(d) Add the following after clause 4.1(n) (a) & (b):-

In case, bidder is a distributor or marketer of a manufacturer or more than one manufacturer, the latest Sales Tax Clearance certificate/returns and the latest Income tax assessment orders/returns of bidder as well as of their manufacturer/manufacturers are to be uploaded.

III. Clause 5. PRICE BID - COVER "B"

FOR:-

5.1. Cover "B" contains the Price Bid of the Tenderer.

(i) (a) The Tenderer shall fill in the rate for quoted item, Central Excise Duty rate in % (percent), rate of CST against form C in % and Total rate (landed price) of quoted item inclusive of Central Excise Duty & CST with form 'C' in respective column of for the items quoted in BOQ.

(ii) **Determination of L1 bidder:** In determining the lowest evaluated price inclusive of Excise duty and CST against form C, the rate quoted per unit landed price as indicated in column No. 9 of the **BOQ** shall be taken into consideration.

iii) The rate quoted in column 9 of **BOQ** should be for a unit and for the given specification. **The rates quoted in paisa are to be in 2 digits.** The Tenderer is not permitted to change/alter specification or unit size given in the **ANNEXURE- IX**

READ:-

5.1. Cover "B" contains the Price Bid of the Tenderer.

i) (a) The Tenderer shall fill in the rate for quoted item, Central Excise Duty rate in % (percent), rate of CST against form C in % and Total rate (landed price) of quoted item inclusive of Central Excise Duty & CST with form 'C' in respective column of for the items quoted in BOQ.

(ii) **Determination of L1 bidder:** In determining the lowest evaluated price inclusive of Excise duty and CST against form C, the rate quoted per unit landed price as indicated in column No. 10 of the **BOQ** shall be taken into consideration.

iii) The rate quoted in column 10 of **BOQ** should be for a unit and for the given specification. **The rates quoted in paisa are to be in 2 digits.** The Tenderer is not permitted to change/alter specification or unit size given in the **ANNEXURE- IX**

IV. Clause 9. ACCEPTANCE OF TENDER

FOR:-

9.1. Evaluation of the tender and determination of the L1 rate (Lowest rate) will be done on the basis of rate per unit landed price as mentioned in column 7 of **BOQ**. Negotiation if required will be done at our premises and the same will be done strictly as per Central Vigilance Commission guidelines.

READ:-

9.1. Evaluation of the tender and determination of the L1 rate (Lowest rate) will be done on the basis of rate per unit landed price as mentioned in column 10 of **BOQ**. Negotiation if required will be done at our premises and the same will be done strictly as per Central Vigilance Commission guidelines.

V. ANNEXURE VIII (CHECK LIST)
Add the following after S.N. 22

23. Copy of sale license in case of bid from the distributor or marketer.

VI. Name and description of Items in ANNEXURE IX , ANNEXURE XI and in BOQ of various Item coded are amended as under: -

FOR:-

| Item Code | Name and description of Items |
|-----------|---|
| 5010 | Sterile Hypodermic Syringe with Needle attached, 24G , Single use-2ml, IS 12050 [Syringe 2 ml with needle 24 G (Sterile, single use, disposable, as per BIS)] |
| 5011 | Sterile Hypodermic Syringe with Needle attached, 24G , Single use-5ml, IS 12050 [Syringe 5 ml with needle 24G (Sterile, single use, disposable, as per BIS)] |
| 5012 | Sterile Hypodermic Syringe with Needle attached, 21G , Single use-10ml, [Syringe 10 ml with needle 21G (Sterile, single use, disposable, as per BIS)] |
| 5013 | Sterile Hypodermic Syringe with Needle attached, 23G , Single use-20ml, [Syringe 20 ml with needle 23G (Sterile, single use, disposable, as per BIS)] |
| 5014 | Needle 16G (Sterile, single use, disposable, as per BIS) |
| 5015 | Needle 18G (Sterile, single use, disposable, as per BIS) |
| 5016 | Needle 23G (Sterile, single use, disposable, as per BIS) |
| 5017 | Needle 26G X ½" (1.27cm)(Sterile, single use, disposable, as per BIS) |
| 6005 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Extra Small Conforms to IS 15354 |
| 6006 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Small Conforms to IS 15354 |
| 6007 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Medium Conforms to IS 15354 |
| 6008 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Large Conforms to IS 15354 |
| 6033 | Insulin syringe (40 units) with (fixed) 30 G needle conform to IS 12227 |
| 7002 | Clinical Thermometer As per IS 3055 (Part-1) Specification, GLASS |
| 8113 | Hernia Belt - made from good quality surgical Elastic with elasticated straps, pads can be adjusted. Size Universal |

READ :-

| Item Code | Name and description of Items |
|-----------|--|
| 5010 | Sterile Hypodermic Syringe with Needle attached, 24G , Single use-2ml, IS 12050 [Syringe 2 ml with needle 24 G (Sterile, single use, disposable, as per BIS/CE Certified)] |
| 5011 | Sterile Hypodermic Syringe with Needle attached, 24G , Single use-5ml, IS 12050 [Syringe 5 ml with needle 24G (Sterile, single use, disposable, as per BIS/CE Certified)] |
| 5012 | Sterile Hypodermic Syringe with Needle attached, 21G , Single use-10ml, [Syringe 10 ml with needle 21G (Sterile, single use, disposable, as per BIS/CE Certified)] |

| | |
|------|--|
| 5013 | Sterile Hypodermic Syringe with Needle attached, 23G , Single use-20ml, [Syringe 20 ml with needle 23G (Sterile, single use, disposable, as per BIS/CE Certified)] |
| 5014 | Needle 16G (Sterile, single use, disposable, as per BIS/CE Certified) |
| 5015 | Needle 18G (Sterile, single use, disposable, as per BIS/CE Certified) |
| 5016 | Needle 23G (Sterile, single use, disposable, as per BIS/CE Certified) |
| 5017 | Needle 26G X ½" (1.27cm)(Sterile, single use, disposable, as per BIS/CE Certified) |
| 6005 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Extra Small Conforms to IS 15354/CE Certified |
| 6006 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Small Conforms to IS 15354/CE Certified |
| 6007 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Medium Conforms to IS 15354/CE Certified |
| 6008 | Rubber examination gloves made of natural rubber latex. Non-sterile,Size: Large Conforms to IS 15354/CE Certified |
| 6033 | Insulin syringe (40 units/100 Units) with (fixed) 30 G needle conform to IS 12227/CE certified. |
| 7002 | Clinical Thermometer As per IS 3055 (Part-1) Specification/CE certified, GLASS |
| 8113 | Hernia Belt - made from good quality surgical Elastic with elasticated straps, pads can be adjusted. Size Universal (SIZE : Small/Medium/Large) |

VII. Name and description of Items, unit size, pack size and tender Quantity in Unit Size in ANNEXURE IX , ANNEXURE XI and in BOQ of various Item Codes are amended as under:-

FOR :-

| Sr. No. | Item Code | Name and description of Items | Unit Size | Pack size | Tender Quantity in Unit Size |
|---------|-----------|---|-----------|-----------|------------------------------|
| 42 | 5077 | Malaria diagnostic tester P.f./pan | 1's | 1'sX10 | 10,000 |
| 43 | 5078 | dengue diagnostic strip | 1's | 1'sX10 | 10,000 |
| 44 | 5079 | Pregnancy diagnostic tester | 1's | 1'sX10 | 10,000 |
| 87 | 6034 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 16G | 1's | One | 30,000 |
| 88 | 6035 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 18G | 1's | One | 300,000 |
| 89 | 6036 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 20G | 1's | One | 300,000 |
| 90 | 6037 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 22G) | 1's | One | 200,000 |
| 91 | 6038 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated . Size 24G suitable for paediatric & neonatal use | 1's | One | 100,000 |

| | | | | | |
|-----|------|---|------|--------------------|---------|
| 99 | 7003 | Sanitary Napkins Regular with wings as Per IS Specification No.5405:1980 | PACK | 6 napkins/ pack | 100,000 |
| 100 | 7004 | Adult Diapers Medium | 1's | One | 50000 |
| 107 | 8083 | a--Sanitary Napkin, Beltless 1. Covering – Covering of the absorbent filler shall be good quality knitted sleeve or non-woven fabric which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. The napkins shall have a non absorbent barrier on one side which shall have an identifying mark indicating the side of the barrier.2. Absorbent Filler – The filler material shall consist of cellulose pulp/ wadding, and shall be free from lumps, oil spots, dirt or foreign material, etc.3. Back Strip – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material.4. Size - The size of absorbent section / complete sanitary napkin shall be as follows: (in mm) Absorbent section Total Pad Length 210 +_ 10 230 +_ 10Width 60 to 75 70 to 85Thickness 8 +_ 25. Weight : Not more than 10 gm | PACK | 6 napkins/ pack | 100,000 |
| 108 | 8084 | b--Sanitary Napkin, Belt-type 1. Covering – Covering of the absorbent filler shall be good quality knitted sleeve or non-woven fabric which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. The napkins shall have a non absorbent barrier on one side which shall have an identifying mark indicating the side of the barrier. 2. Absorbent Filler – The filler material shall consist of cellulose pulp/ wadding, and shall be free from lumps, oil spots, dirt or foreign material, etc. 3. Size - The size of absorbent section / complete sanitary napkin shall be as follows: (in mm) Absorbent section Pad Length 220+_10 ; Width 70+_5; Thickness 17+_3 4. Weight : 12 +_ 3 gm 5. Pack – Six napkins in a pack. 6. Elastic Belt with loops shall be provided in each pack. 7. Absorbency: The napkin should be able to absorb not less than 30 ml of normal saline or coloured water or test fluid when poured on to the centre of the napkin at the rate of 15 ml per minute. | PACK | 6 napkins/ pack | 100,000 |

READ :-

| Sr. No. | Item Code | Name and description of Items | Unit Size | Pack size | Tender Quantity in Unit Size |
|---------|-----------|---|-------------------|-----------------|------------------------------|
| 42 | 5077 | Malaria Antigen Test Kit for all type of malarial infection | 1's | 1'sX10 | 100,000 |
| 43 | 5078 | Dengue Antigen (NS1) Test Kit | 1's | 1'sX10 | 50,000 |
| 44 | 5079 | HCG -Pregnancy Card Test Kit | 1's | 1'sX10 | 100,000 |
| 87 | 6034 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 16G | 1's | 1'sx50 | 30,000 |
| 88 | 6035 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 18G | 1's | 1'sx50 | 300,000 |
| 89 | 6036 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 20G | 1's | 1'sx50 | 300,000 |
| 90 | 6037 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated 3 Way stop cock. Size 22G) | 1's | 1'sx50 | 200,000 |
| 91 | 6038 | Sterile Disposable (Single Use) Teflon/ PTFE I.V. Cannula with integrated . Size 24G suitable for paediatric & neonatal use | 1's | 1'sx50 | 100,000 |
| 99 | 7003 | Napkin Regular with wings as per size pad length 200±20mm, width : 60 to 70mm and thickness 15±2mm and remaining parameters confirming to specification no. 5405 : 1980/CE Certified. | PACK of 6 napkins | 6 napkins/ pack | 100,000 |
| 100 | 7004 | Adult Diapers Medium | 1's | One | 50000 |
| 107 | 8083 | a--Sanitary Napkin, Beltless 1. Covering – Covering of the absorbent filler shall be good quality knitted sleeve or non-woven fabric which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. The napkins shall have a non absorbent barrier on one side which shall have an identifying mark indicating the side of the barrier. 2. Absorbent Filler – The filler material shall consist of cellulose pulp/ wadding, and shall be free from lumps, oil spots, dirt or foreign material, etc. 3. Back Strip – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4. Size - The size of absorbent section / complete sanitary napkin shall be as follows: (in mm) | PACK of 6 napkins | 6 napkins/ pack | 100,000 |

| | | | | | |
|-----|------|--|-------------------|-----------------|---------|
| | | Absorbent section Total Pad Length 210±10, 230 ± 10 Width 60 to 75, 70 to 85, Thickness 8 ± 2 5. Weight : Not more than 10 gm | | | |
| 108 | 8084 | b--Sanitary Napkin, Belt-type 1. Covering – Covering of the absorbent filler shall be good quality knitted sleeve or non-woven fabric which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. The napkins shall have a non absorbent barrier on one side which shall have an identifying mark indicating the side of the barrier. 2. Absorbent Filler – The filler material shall consist of cellulose pulp/ wadding, and shall be free from lumps, oil spots, dirt or foreign material, etc. 3. Size - The size of absorbent section / complete sanitary napkin shall be as follows: (in mm) Absorbent section Pad Length 220 ± 10 Width 70±5 , Thickness 17±3 4. Weight : 12 ± 3 gm 5. Pack – Six napkins in a pack. 6. Elastic Belt with loops shall be provided in each pack. 7. Absorbency: The napkin should be able to absorb not less than 30 ml of normal saline or coloured water or test fluid when poured on to the centre of the napkin at the rate of 15 ml per minute. | PACK of 6 napkins | 6 napkins/ pack | 100,000 |

Note:- The description, unit size and quantity of item code 7003 and 7004 mentioned in BOQ is not correct and the same may read as above.

VIII. Bidders are requested to quote their rates in BOQ considering the modified name, unit Size , pack size, quantity as amended above.

All other contents of the tender document are unaltered.

(Mahadev Agarwal)
Manager(Procurement)