

Amendment no. 1

No.: - PMBI/SURGICAL/RC-167/2021

Dated: 28.12.2021

Subject: - Tender No. PMBI/SURGICAL/RC-167/2021 dated 10/12/2021 for supply of OXO-BIODEGRADABLE SANITARY NAPKINS, DIAPERS, GLOVES & OTHER CONSUMABLE ITEMS (12 items) to Pharmaceuticals and Medical Devices Bureau of India (PMBI).

Reference: - Pre-Bid Meeting held on 17/12/2021 at 11:00 AM in the premises of PMBI.

Pharmaceuticals and Medical Devices Bureau of India (PMBI) has invited e-Bids from the interested parties for “e- Tender for Supply of Oxo-Biodegradable Sanitary Napkins, Diapers, Gloves & Other Consumable Items for the year 2021- 2023”, vide Notice Inviting e-Tender No.- **PMBI/SURGICAL/RC-167/2021**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website of Central Public Procurement Portal; www.eprocure.gov.in and BPPI Website; www.janaushadhi.gov.in.

After considering the suggestions/queries received from the prospective bidders on or before Pre-Bid meeting, the clarifications/amendments in respect to the tender document and specification of the tendered items have been made as per below Annexure-A respectively. All other technical specifications, terms, and conditions along with the tender schedule as mentioned in tender document shall remain unchanged.

The following amendment in Tender Document is hereby authorized: -

Annexure- A

Sl. No.	Tender Clause / Reference	Query / Suggestion	Clarification / Amendment
1.	5. PRICE BID – “COVER-B” (Financial Bid/BOQ)	Contract no. in uploaded BOQ in macro excel sheet kept as PMBI/LTD./DRUG-167/2021	The bidders must consider the contract no. as PMBI/SURGICAL/RC-167/2021 in the BOQ while quoting the rates.
2.	Clause no. 3. K. ELIGIBILITY CRITERIA (TECHNICAL BID - COVER “A”):	One of the bidders has requested to increase the turnover criteria of last four consecutive financial years from 2 (two) Crores to higher showing the tender value.	Tender condition prevails.
3.	Clause no. 18.M, Annexure - XII Secondary Pack	Secondary Pack size against Item Code 8140 is mentioned as Pack of 4's x 10.	Secondary Pack size against Item Code 8140 is hereby amended as Pack of 4's x 100.
4.	Clause no. 3. N. ELIGIBILITY CRITERIA (TECHNICAL BID - COVER “A”):	One of the Bidders has requested to relax the Market standing certificate for start-ups.	Tender condition prevails.

5.	Clause no. 19. N. SUPPLY CONDITIONS:	One of the Bidders has asked to consider the Pandemic and force majeure events equally. They have also requested to increase the scheduled delivery period from 45 days to 60-90 days stating the disruption in supply chain management of raw materials of the tendered items.	Tender condition prevails.
6.	Clause no. 21 PACKING:	Bidders shall provide a pouch or wrapper for disposal of each napkin or diapers along with the packet of their sanitary products.	Tender condition prevails.
7.	Clause 18 (M) Annexure XII	<p>One of the bidders has raised a query regarding specification and packing of the Drug Code 6006</p> <p>Rubber Examination Gloves, Non-sterile (Small) Rubber examination gloves made of natural rubber latex. Made of natural rubber Latex, powder free, without tear, properly folded in a paper, Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA.</p>	<p>The specification is amended as follows: Rubber Examination Gloves, Non-sterile (Small)</p> <p>Rubber examination gloves made of natural rubber latex. powder free, without tear, Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2. Should Conform to IS 13422 ISI marked / CE certified / FDA.</p> <p>Unit Size is 1's 50 gloves should be packed in 1 Box i.e., 1's X 50 (Primary packing) The outer carton packing should be as per Annexure VIII point (ii).</p>
8.	Clause 18 (M) Annexure XII	<p>One of the bidders has raised a query regarding specification and packing of the Drug Code 6007</p> <p>Rubber Examination Gloves, Non-sterile (Medium) Rubber examination gloves made of natural rubber latex. Made of natural rubber Latex, powder free, without tear, properly folded in a paper, Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic should Conform to IS 13422 ISI marked / CE certified / FDA</p>	<p>The specification is amended as follows: Rubber Examination Gloves, Non-sterile (Medium)</p> <p>Rubber examination gloves made of natural rubber latex, powder free, without tear, Free of holes, with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2. Should Conform to IS 13422 ISI marked / CE certified / FDA.</p> <p>Unit Size is 1's 50 gloves should be packed in 1 Box i.e., 1's X 50 (Primary packing) The outer carton packing should be as per Annexure VIII point (ii)</p>

9.	Clause 3. E (e)	Reports of Oxo-biodegradable material used	All the reports as required under Clause 3.E (e) for Oxo-biodegradable Sanitary napkins shall be furnished with the technical documents.
10.	Annexure I, Checklist	Additional point of Annexure I, Checklist	Additional point to be added in the Annexure I, Checklist as Sr. No. 27 Supporting documents /reports /certificate to prove biodegradability of 100 % Oxo-biodegradable material as per tender clause 3.E (e).
11.	Clause 3.J	Tenderer must declare their maximum Production Capacity (item wise) issued by concerned Licensing Authority / self-declaration highlighting the quoted product as per Clause no. 3. J.	Tenderer must declare their maximum Production Capacity (item wise) issued by concerned Licensing Authority / CA certified highlighting the quoted product as per Clause no. 3. J.

All other contents of tender document remain unaltered. Bidders are requested to quote their rates considering all the terms and condition of the tender document including Amendment no. 1 dated 28.12.2021.

Sd/-
DGM (Procurement & Quality)
For & on behalf of PMBI
Ph: 011-49431811