

### **Amendment no. 5**

**No.: -BPPI/SURGICAL/RC-158/2020**

**Dated: 14/12/2020**

**Subject: - Tender No. BPPI/SURGICAL/RC-158/2020 Dated 06/11/2020 for procurement of Surgical/ Consumables and Medical Devices to Bureau of Pharma Public Sector undertakings of India (BPPI).**

**Reference: - Amendment No. 2 dated 25/11/2020**

Bureau of Pharma PSUs of India (BPPI) has invited e-Bids from the interested parties for “e-Tender for the procurement of Surgical/ Consumables and Medical Devices for the year 2020- 2022”, vide Notice Inviting Tender No.- BPPI/SURGICAL/RC-158/2020. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website of Central Public Procurement Portal; [www.eprocure.gov.in](http://www.eprocure.gov.in) and BPPI Website; [www.janaushadhi.gov.in](http://www.janaushadhi.gov.in).

The following amendments regarding the specification in tender document is hereby authorized as below: -

<b>Sl. No.</b>	<b>Tender Clause/ Reference</b>	<b>Item Code with Generic Name and specification of the item</b>	<b>Amendment</b>
1	Clause 18 (M) Annexure -XII, XIII & BOQ	<b>Item code-8121:</b>  GLUCOMETER TEST STRIP  Note: Item code 8121 shall be awarded with item code 8122 "Glucometer Digital" and should not be awarded individually. Hence, both items i.e., item code 8121 & 8122 must be qualified	<b>Item code-8121:</b>  Glucometer Test Strip (Compatible with Janaushadhi Glucometer Digital)
2	Clause 18 (M) Annexure -XII, XIII & BOQ	<b>Item Code 8122.</b> GLUCOMETER DIGITAL (1 glucometer, 25 lancets,1 lancing device,1 battery 3v., Warranty card) <b>plus a pack of 25 strips.</b>  Glucometer Specification: Glucose ranges: 30 - 800 md/dL Analysis time: 5 Sec Sample Volume: < 1.0 micro-litres Sample: Whole Blood Sample Mode: Capillary Fill detection: Automatic	<b>Item Code 8122.</b>  <b>GLUCOMETER DIGITAL (1 glucometer, 25 strips, 25 lancets, 1 lancing device, 1battery 3v. &amp; Warranty card)</b>  <b>Glucometer Specification:</b> Glucose ranges : 30 - 800 md/dL Analysis time : 5 Sec Sample Volume : < 1.0 micro-litres Sample : Whole Blood

	<p>Battery: 3v Lithium Battery (Replaceable) (CR2032), should work for minimum 1000 test  Minimum memory: 50 samples (Erasable),  Replacement Warranty Period: 5 Years,  Code: Auto Coding/No Coding,  Display: 48mm (diagonally) LCD (<math>\pm 5</math>mm).  Strip Ejector facility to avoid contamination with blood.</p> <p>ISO-15197:2013 (International Standard) CE Certificate NIB (National Institute of Biological) Tested.</p> <p>Customer Care Toll Free Number: Customer Care Toll Free Number must be provided by the Awardee bidder for the customer support.</p> <p>Note: i) A pack of 25 strips should be separately placed with the mono-carton pack of Digital Glucometer.</p> <p>Note: ii) Item code 8122 shall be awarded with item code 8121 "Glucometer Test strips" and should not be awarded individually. Hence, both items i.e. item code 8121 &amp; 8122 must be qualified.</p>	<p>Sample Mode : Capillary  Fill detection : Automatic  Battery : 3v Lithium Battery (Replaceable) (CR2032), should work for minimum 1000 test  Minimum memory : 50 samples (Erasable),  Replacement Warranty Period: 3 Years,  Code : Auto Coding/No Coding,  Display : 48mm (diagonally) LCD (<math>\pm 5</math>mm).  Strip Ejector facility to avoid contamination with blood.</p> <p>ISO-15197:2013 (International Standard), CE Certificate, NIB (National Institute of Biological) Tested.</p> <p><b>Note:</b> Control Solution for calibration of device shall be provided.</p> <p><b>Customer Care Toll Free Number:</b> Customer Care Toll Free Number must be provided by the Awardee bidder for the customer support.</p>
--	---	--

All other contents of tender document remain unaltered. Bidders are requested to quote their rates considering all the terms and condition of the tender document including all Amendments published against subject tender.

Yours faithfully,

SD/-

(Manik Bera)  
Deputy Manager (Procurement)  
For & on behalf of BPPI  
Ph: 011-49431854