

**PHARMACEUTICALS AND MEDICAL DEVICES BUREAU OF INDIA (PMBI)**  
**(Set up under the Department of Pharmaceuticals, Govt. of India)**  
8<sup>th</sup> Floor, Videocon Tower, Block E1, Jhandewalan Extension, New Delhi-110055

**Amendment No. 2**

**Ref. Tender No: - PMBI/SURGICAL/RC-201/2022**

**Dated: 05/08/2022**

**Subject: - Tender No. PMBI/SURGICAL/RC-201/2022 dated 07/07/2022 for supply of Surgical/ Consumables and Medical Devices to Pharmaceuticals and Medical Devices Bureau of India (PMBI).**

Pharmaceuticals and Medical Devices Bureau of India (PMBI) has invited e-Bids from the interested parties for "e- Tender for Supply of **Surgical/Consumables and Medical Devices** for the year 2022- 2024", vide Notice Inviting e-Tender No.- **PMBI/SURGICAL/RC-201/2022**. Detailed tender document containing eligibility criteria, selection mechanism, other terms and conditions are available on the website of Central Public Procurement Portal; [www.eprocure.gov.in](http://www.eprocure.gov.in) and PMBI Website; [www.janaushadhi.gov.in](http://www.janaushadhi.gov.in).

After considering the suggestions/queries received from the prospective bidders on or before Pre-Bid meeting, the clarifications/amendments in respect to the tender document and specification of the tendered items have been made as per below mentioned Part-A, Part-B & Part C respectively. All other technical specifications, terms, and conditions along with the tender schedule as mentioned in tender document shall remain unamended. The following amendment in Tender Document is hereby authorized: -

The following amendment in Tender Document is hereby authorized as per Part – A, Part -B and Part - C: -

**Part – A**

The following amendment in the tender is hereby authorized as follow: -

Sl. No.	Reference Clause / Annexure No.	Objection / Clarification	Amendment
1	3.A, Note (i) & (ii) and 6. B & C	Bidder has requested to clarify on exemption of EMD on submission of Udyam Registration Certificate.	All bidders are invited to see tender provision under Clause 3. Note (i) where it is clearly mentioned "Tenderer may be exempted from the payment of EMD, if valid registration certificate from NSIC/MSME/ Udyam Registration Certificate is uploaded" Accordingly, tender clause no. 6.B is amended as under: "Tenderer may be exempted from the payment of EMD, if valid registration certificate from NSIC/MSME/Udyam Registration Certificate is uploaded issued in the name of firm for which bidder has submitted quotation".

	3. E. (c)	Bidders must submit test certificates/reports for all contraceptive items in the tender in support of its Biocompatibility conforming ISO 10993 along with other standards/reports conforming ISO 12243, ISO 29941 etc. items wherever applicable. Packaging shall confirm ISO 25841:2017.	The clause is modified as under: - Bidders must submit test certificates/reports for all contraceptive items in the tender in support of its Biocompatibility conforming ISO 10993 along with other standards/reports conforming ISO 12243, ISO 29941 etc. items wherever applicable. For Male and Female condom Specification and packaging shall confirm ISO 4074:2015 and ISO 25841:2017 respectively .
2	3.F	Bidder has requested to clarify on submission of Market Standing Certificate for medical devices which has recently been covered under Medical Devices.	It is to clarify intending bidders that in case the items have recently been covered under Medical Devices and Licensing Authority denies issuance of Market Standing Certificate, bidder may submit a declaration with Market standing Certificate (MSC) issued by the C.A./C.S certifying at least three batch No. of the quoted items that the firm/company has manufactured and marketed the items for last three consecutive years.
3	4.P	Some of the intending bidder has raised query during pre-bid meeting to clarify whether reference samples of quoted item shall be in their company artwork.	It is to clarify that bidder may submit reference samples of quoted item in their company/trade artwork.
4	3.D & E	One of the bidders has requested to clarify the requirement of manufacturing license in case if bidder participate against Menstrual Cup saying the item is not regulated under Medical Device as of now.	It is to clarify the intending bidders that in case Menstrual Cup is not regulated as Medical Device as of now, intending bidder may submit their product manufacturing permission/registration along with a declaration on non-judicial stamp paper highlighting 'the item is not covered under Medical Devices as of now and they promise to submit all document approved by Licensing Authority if during contract period the product gets regulated as medical devices.
5	19. D	One of the bidders has requested to increase delivery period from 45 days vide their mail.	Tender condition prevails.
6	23	One of the bidders has requested to reduce payment from 60 Days to 45 days.	Tender condition prevails.
7	<b>3.A, and 6.A for format of Model Bank Guarantee (BG) for furnishing EMD and Clause no. 3.U for 'DECLARATION OF LOCAL CONTENT'</b>	The format of Model Bank Guarantee (BG) for furnishing EMD has been misplaced under Annexure III of the tender document on page no. 38/75.	The format of Model Bank Guarantee (BG) for furnishing EMD shall be read under Annexure X referring Clause no. 3.A, and 6.A of the tender documents on page no. 53/75 while format for 'DECLARATION OF LOCAL CONTENT' under Annexure X shall be considered under Annexure- XVI on page 73/75 only referring Clause no. 3.U.
8	4.P (iv) (Sample)	iv) Firm may take back their samples if unapproved within 10 days from the date of issue of the Rate	Modification is hereby authorized: - iv) Firm may take back their samples if unapproved/bids are technically disqualified within 10 days from the date of technical

		contract, otherwise the same will be destroyed by PMBI.	disqualification on CPP portal, otherwise the same will be destroyed by PMBI. Any claim made after 10 days shall not be considered.
9	4.P, Annexure XVIII	Heading under Annexure XVII 'Please receive following Samples of Oxo-Biodegradable Sanitary Napkins, Diapers& Other Consumable Items from'.	The term ' <b>Samples of Oxo-Biodegradable Sanitary Napkins, Diapers&amp; Other Consumable Items from</b> ' is hereby replaced with ' <b>Samples of Surgical/ Consumables and Medical Devices</b> '. <b>Read Annexure XVII as Annexure XVIII.</b>
10	<b>Clause 18 (M), Annexure -XII</b>	For Item code 9003, Secondary pack is mentioned as One in Mono-Pack X 10	Secondary pack shall be read as Three in Mono-Pack X 200

### Part- B

The following amendment is hereby authorized in Specification of the items and unit size:

(1)	(2)	(3)	(4)	(5)
S.N.	Item Code	Generic Name of Item	Amendment in the specification	Unit Size
1	8133	<b>Digital Blood Pressure Instrument (100% mercury free):</b> Fully Automated upper arm style with LCD Digital Display/Digital monitor with acrylic panel and 3 colour backlit display for indication on display, Display shall show Numerical Data for reading. Displayed parameters: Unit should display systolic pressure, diastolic pressure, pulse rate and mean arterial pressure. Dimension: Design shall be handy and Portable and as per market standard. Measuring Method: Oscillometric technique. Measurement Range: Pressure- 0mmHg – 300 mmHg (0.00 kPa to 39.9 kPa), Pulse- 30 to 199 Beats/Minute Measurement Accuracy: Pressure- $\pm 3$ mmHg Pulse- $\pm 5\%$ , Measurement time (s): $\leq 60$ , user selectable. Power supplies: AA / AAA Alkaline Battery and USB operated. Memory: 2 x 99 Memory (90-99 memory). Storage conditions: Temperature range when in use: from 5°C to 40°C. 15% to 90% RH. Storage and transport conditions: - 20°C to 60°C. 10% to 95% RH. Equipment alarms preferred: Hose leak, inflation, or deflation error, Irregular Heartbeat Detection, Low Battery alarm, reading failure alarm. Accessories: High quality Arm Cuff (dimension 22 cm to 32 cm) with tube of suitable length, Protective case. USB Cable, 4 AA / AAA Alkaline battery set, User manual, Warranty Card. Shelf Life and warranty: Five Years with 2-year warranty (from the date of sale at Janaushadhi Kendra) Clinically validated algorithm International Standards: ISO 13485:2016 Certified, FDA Certified, CE certified.	<b>Digital Blood Pressure Instrument (100% mercury free):</b> Fully Automated upper arm style with LCD Digital Display/Digital monitor with WHO indicator and acrylic panel, Display shall show Numerical Data for reading. Displayed parameters: Unit should display systolic pressure, diastolic pressure, pulse rate Dimension: Design shall be handy and Portable and as per market standard. Display size shall be : 4.5-5.5 cm X 5 – 6 cm ( $\pm 1$ cm) Measuring Method: Oscillometric technique. Measurement Range: Pressure- 0mmHg – 300 mmHg (0.00 kPa to 39.9 kPa), Pulse- 30 to 199 Beats/Minute Measurement Accuracy: Pressure- $\pm 3$ mmHg Pulse- $\pm 5\%$ , Measurement time (s): $\leq 60$ , user selectable. Power supplies: AA / AAA Alkaline Battery and Micro/ C-type/Round Pin USB operated. Memory: 2 x 99 Memory s(90-99 memory). Storage conditions: Temperature range when in use: from 5°C to 40°C. 15% to 90% RH. Storage and transport conditions: - 20°C to 60°C. 10% to 95% RH. Equipment alarms preferred: Hose leak, inflation, or deflation error, Irregular Heartbeat Detection, Low Battery alarm, reading failure alarm. Accessories: High quality Arm Cuff (dimension 22 cm to 32 cm) with tube of suitable length, Protective case. USB Cable type-C (100cm), 4 AA / AAA Alkaline battery set, User manual, Warranty Card. Shelf Life and warranty: Five Years with 2-year warranty (from the date of sale at Janaushadhi Kendra) Clinically validated algorithm International Standards: ISO 13485:2016 Certified/ FDA Certified/ CE certified.	One in Mono-Pack
2	8136	<b>Electric Nebulizer Machine:</b> Compressor Nebulizer Machine Motor: Piston Compressor Motor with 25mm Copper binding Body: ABS body	<b>Electric Nebulizer Machine:</b> Compressor Nebulizer Machine Motor: Piston Compressor Motor with 25mm Copper coated Aluminium binding. Body: ABS body	One in Mono-Pack

		<p>Particle Size: &lt; 0.5 um (µm)  Sound: 50-55 dB  Standards and Safety Certificates: ISO 13485:2016 Certified, WHO-GMP Certified, FDA / CE; ISO 27427-2010.  Warranty: 2 Years  Weight (lbs, kg) / Design: &lt;2kg and design shall be handy and Portable.  Medication Cup: Shall be scaled for 2-6 ml medication  Power Consumption: Product shall be power efficient  Knob/Switch: Good quality indicator switch.  Storage conditions: Temperature range when in use: from 5°C to 40°C. 15% to 90% RH.  Storage and transport conditions: - 20°C to 60°C. 10% to 95% RH.  Electric Cable: ISI certified 1.5 meter long, Plug shall be so designed to save from any shock while plugging in.  Accessories: Compressor Nebulizer Machine, Adult mask, Child Mask, Mouthpiece, Medicine Cup, Filter (8pcs), Air Tube, Warranty Card, and Pouch.  Nebulizer Mask with Tubing: shall be free from any kind of odour, made of clear, Non-toxic PVC, medium concentration, adjustable nose clip &amp; non-autoclave Tube Length: 2-meter,  Nebulization rate: 3cc / 10 mins.  Packing: All the Accessories shall be supplied with good quality storage pouch (Made of clothing/durable material).</p>	<p>Particle Size: &lt; 0.5 um (µm)  Sound: 50-55 dB  Standards and Safety Certificates: ISO 13485:2016 Certified, WHO-GMP Certified, FDA / CE; ISO 27427-2010.  Warranty: 2 Years  Weight (lbs, kg) / Design: &lt;1.5 kg and design shall be handy and Portable.  Medication Cup: Shall be scaled for 2-6 ml medication  Power Consumption: Product shall be power efficient  Knob/Switch: Good quality indicator switch.  Storage conditions: Temperature range when in use: from 5°C to 40°C. 15% to 90% RH.  Storage and transport conditions: - 20°C to 60°C. 10% to 95% RH.  Electric Cable: ISI certified 1.5 meter long, Plug shall be so designed to save from any shock while plugging in.  Accessories: Compressor Nebulizer Machine, Adult mask, Child Mask, Mouthpiece, Medicine Cup, Filter (8pcs), Air Tube, Warranty Card, and Pouch.  Nebulizer Mask with Tubing: shall be free from any kind of odour, made of clear, Non-toxic PVC, medium concentration, adjustable nose clip &amp; non-autoclave Tube Length: 2-meter,  Nebulization rate: 3cc / 10 mins.  Packing: All the Accessories shall be supplied with good quality storage pouch (Made of clothing/durable material).</p>	
3	9000	<p><b>Reusable Menstrual Cup (non-absorbent bell-shaped hygienic device)-Small:</b>  Medical grade silicone shall be used with medical grade platinum catalyst for polymer cross linkage, FDA/CE Compliance, Shall be Biocompatible.  Air holes: At least two air hole of 1 mm size close to the rim covering 2 side.  Cup Wall thickness: 2 mm (± 0.3 mm)  Pull out stem length: minimum 15 cm  Size: Small; Diameter of Outer rim: 36-40 mm  Length of the cup excluding pull  Out stem: approx. 40-50mm  Cup Storage capacity: 15-25 ml  Minimum Shelf life: Minimum 5 Years as per ISO-2230:2002  Colour: Colourless/Transparent  Testing: Cup shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization).  Regulation and Conformity: ISO 9001:2015 or ISO 13485 certified  *( shall be supplied with protective textile pouch and user instructions.)  Material shall be biocompatible ensuring it no change in performance after multiple uses, good tear strength and endures boiling, cyclic use and pulling out, easy insertion and removal, dimensionally stable.</p>	<p><b>Reusable Menstrual Cup (non-absorbent bell-shaped hygienic device)-Small:</b>  Medical grade silicone shall be used with medical grade platinum catalyst for polymer cross linkage, FDA/CE Compliance, Shall be Biocompatible.  Air holes: At least two air hole of 1 mm size close to the rim covering 2 side.  Cup Wall thickness: 2 mm (± 0.3 mm)  Pull out stem length: minimum 15 mm ± 1 mm  Size: Small; Diameter of Outer rim: 36-40 mm  Length of the cup excluding pull  Out stem: approx. 40-50mm  Cup Storage capacity: 15-25 ml  Minimum Shelf life: Minimum 5 Years as per ISO-2230:2002  Colour: Colourless/Transparent  Testing: shall comply ISO 10993-5, 10, 23 (if any other applicable).  Regulation and Conformity: ISO 9001:2015 or ISO 13485 certified  *( shall be supplied with good quality protective textile pouch and user instructions.)  Material shall be biocompatible ensuring it no change in performance after multiple uses, good tear strength and endures boiling, cyclic use and pulling out, easy insertion and removal, dimensionally stable.</p>	One in Mono-Pack
4	9001	<p><b>Reusable Menstrual Cup (non-absorbent bell-shaped hygienic device)-Medium:</b>  Medical grade silicone shall be used with medical grade platinum catalyst for polymer cross linkage, FDA/CE Compliance, Shall be Biocompatible,  Air holes: At least two air hole of 1 mm size close to the rim covering 2 side.  Cup Wall thickness: 2 mm (± 0.3 mm)  Pull out stem length: minimum 15 cm  Size: Small; Diameter of Outer rim: 41-44 mm</p>	<p><b>Reusable Menstrual Cup (non-absorbent bell-shaped hygienic device)-Medium:</b>  Medical grade silicone shall be used with medical grade platinum catalyst for polymer cross linkage, FDA/CE Compliance, Shall be Biocompatible,  Air holes: At least two air hole of 1 mm size close to the rim covering 2 side.  Cup Wall thickness: 2 mm (± 0.3 mm)  Pull out stem length: minimum 15 mm ± 1 mm  Size: Small; Diameter of Outer rim: 41-44 mm</p>	One in Mono-Pack

		<p>Length of the cup excluding pull Out stem: approx. 45-55mm</p> <p>Cup Storage capacity: 20-30 ml</p> <p>Minimum Shelf life: Minimum 5 Years as per ISO-2230:2002</p> <p>Colour: Colourless/Transparent</p> <p>Testing: Cup shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization).</p> <p>Regulation and Conformity: ISO 9001:2015 or ISO 13485 certified</p> <p>* (shall be supplied with protective textile pouch and user instructions).</p> <p>Material shall be biocompatible ensuring it no change in performance after multiple uses, good tear strength and endures boiling, cyclic use and pulling out, easy insertion and removal, dimensionally stable.</p>	<p>Length of the cup excluding pull Out stem: approx. 45-55mm</p> <p>Cup Storage capacity: 20-30 ml</p> <p>Minimum Shelf life: Minimum 5 Years as per ISO-2230:2002</p> <p>Colour: Colourless/Transparent</p> <p>Testing: shall comply ISO 10993-5, 10, 23 (if any other applicable).</p> <p>Regulation and Conformity: ISO 9001:2015 or ISO 13485 certified.</p> <p>* (shall be supplied with good quality protective textile pouch and user instructions).</p> <p>Material shall be biocompatible ensuring it no change in performance after multiple uses, good tear strength and endures boiling, cyclic use and pulling out, easy insertion and removal, dimensionally stable.</p>	
5	9002	<p><b>Reusable Menstrual Cup (non-absorbent bell-shaped hygienic device)-Large:</b></p> <p>Medical grade silicone shall be used with medical grade platinum catalyst for polymer cross linkage, FDA/CE Compliance, Shall be Biocompatible,</p> <p>Air holes: At least two air hole of 1 mm size close to the rim covering 2 side.</p> <p>Cup Wall thickness: 2 mm (<math>\pm 0.3</math> mm)</p> <p>Pull out stem length: minimum 15 cm</p> <p>Size: Small; Diameter of Outer rim: 45-48 mm</p> <p>Length of the cup excluding pull Out stem: approx. 48-58mm</p> <p>Cup Storage capacity: 30-40 ml</p> <p>Minimum Shelf life: Minimum 5 Years as per ISO-2230:2002</p> <p>Colour: Colourless/Transparent</p> <p>Testing: Cup shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization).</p> <p>Regulation and Conformity: ISO 9001:2015 or ISO 13485 certified</p> <p>* shall be supplied with protective textile pouch and user instructions).</p> <p>Material shall be biocompatible ensuring it no change in performance after multiple uses, good tear strength and endures boiling, cyclic use and pulling out, easy insertion and removal, dimensionally stable.</p>	<p><b>Reusable Menstrual Cup (non-absorbent bell-shaped hygienic device)-Large:</b></p> <p>Medical grade silicone shall be used with medical grade platinum catalyst for polymer cross linkage, FDA/CE Compliance, Shall be Biocompatible,</p> <p>Air holes: At least two air hole of 1 mm size close to the rim covering 2 side.</p> <p>Cup Wall thickness: 2 mm (<math>\pm 0.3</math> mm)</p> <p>Pull out stem length: minimum 15 mm <math>\pm 1</math> mm</p> <p>Size: Small; Diameter of Outer rim: 45-48 mm</p> <p>Length of the cup excluding pull Out stem: approx. 48-58mm</p> <p>Cup Storage capacity: 30-40 ml</p> <p>Minimum Shelf life: Minimum 5 Years as per ISO-2230:2002</p> <p>Colour: Colourless/Transparent</p> <p>Testing: Testing: shall comply ISO 10993-5, 10, 23 (if any other applicable).</p> <p>Regulation and Conformity: ISO 9001:2015 or ISO 13485 certified</p> <p>* shall be supplied with good quality protective textile pouch and user instructions).</p> <p>Material shall be biocompatible ensuring it no change in performance after multiple uses, good tear strength and endures boiling, cyclic use and pulling out, easy insertion and removal, dimensionally stable.</p>	One in Mono-Pack
6	9003	<p><b>Female Natural Latex Condom:</b></p> <p>Pouch Dimension: Total Device Length: 125-145 mm</p> <p>Width: 66-77 <math>\pm 2</math> mm</p> <p>Thickness: 0.09 - 0.12 mm</p> <p>External Retainer: Made of High-Grade Polyethylene</p> <p>Diameter: 70-80 mm</p> <p>Thickness (in mm): 1.5 – 2.5 mm</p> <p>Width (in mm): 2.0-4.0 mm</p> <p>Colour and Shape: Natural / white and Suitable</p> <p>Internal Retainer: Made of Polyurethane open cell foam</p> <p>Outer diameter (in mm): 47-53 mm</p> <p>Inner diameter (in mm): 7.0 – 13 mm</p> <p>Thickness (in mm): 11.0 – 13.0 mm</p> <p>Colour and Shape: Natural / white and Suitable</p>	<p><b>Female Natural Latex Condom:</b></p> <p>Pouch Dimension: Total Device Length: Minimum 125 mm with tolerance of <math>\pm 10</math> mm</p> <p>Width: 66-77 <math>\pm 2</math> mm</p> <p>Thickness: 0.09 - 0.12 mm</p> <p>External Retainer: Made of High-Grade Polyethylene</p> <p>Diameter: 70-80 mm</p> <p>Thickness (in mm): 1.5 – 2.5 mm</p> <p>Width (in mm): 2.0-4.0 mm</p> <p>Colour and Shape: Natural / white and Suitable</p> <p>Internal Retainer: Made of Polyurethane open cell foam</p> <p>Outer diameter (in mm): 47-53 mm</p> <p>Inner diameter (in mm): 7.0 – 13 mm</p> <p>Thickness (in mm): 11.0 – 13.0 mm</p> <p>Colour and Shape: Natural / white and Suitable</p>	Three in Mono-Pack

		<p>Lubricant: Silicon Oil, minimum 750mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pine Apple / Bubble-gum.  Testing: Condom shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737:2018.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941.  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017  Hazardous Raw Material shall strictly not be Used.  Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall adhere to Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.</p>	<p><i>(if bidder provides ring, they shall have approved design by Authority)</i></p> <p>Specification, Packaging: Shall comply ISO 25841:2017</p> <p>Lubricant: Silicon Oil, minimum 750mg including powder (Calcium carbonate/Magnesium carbonate) not more than 100 mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pine Apple / Bubble-gum.  Testing: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.</p> <p>For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941</p> <p>Conforming QMS 13485:2016/NS-EN ISO 13485:2016</p> <p>Hazardous Raw Material shall strictly not be Used.  Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall adhere to Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.</p>	
7	9004	<p><b>Male Latex Condom (Plain):</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pineapple / Bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941. Product Shall confirm ISO 25841-2017  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	<p><b>Male Latex Condom (Plain): Made from Natural rubber latex</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg including powder (Calcium carbonate/Magnesium carbonate) not more than 50mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Pan / Rose / Guava / Orange / Pineapple / Bubble-gum.</p> <p>Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.  <b>For Microbial content product shall comply</b> ISO 11737 / USP 61 &amp; 62.  For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941.  Specification and packaging shall comply BIS/ISO 4074:2015 and shall be square shaped minimum 55mm x 55mm  Conforming ISO 13485:2016/NS-EN ISO 13485:2016</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall pre-qualified be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	Three in Mono-Pack

8	9005	<p><b>Male Latex Condom (Extra Thin):</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.040 ± 0.060 mm  Lubricant: Silicon Oil; Not less than 350mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pineapple / bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941. Product Shall confirm ISO 25841-2017  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	<p><b>Male Latex Condom (Extra Thin): Made from Natural rubber latex</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.040 - 0.060 mm  Lubricant: Silicon Oil; Not less than 350mg including powder (Calcium carbonate/Magnesium carbonate) not more than 50mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Pan / Rose / Guava / Orange / Pineapple / Bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.  <b>For Microbial content product shall comply</b> ISO 11737 / USP 61 &amp; 62.  For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941.  Specification and packaging shall comply BIS/ISO 4074:2015 and shall be square shaped minimum 55mm x 55mm  Conforming ISO 13485:2016/NS-EN ISO 13485:2016</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall pre-qualified be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	Three in Mono-Pack
9	9006	<p><b>Male Latex Condom (Normal Dot):</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pine Apple / bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941. Product Shall confirm ISO 25841-2017  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	<p><b>Male Latex Condom (Normal Dot): Made from Natural rubber latex</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg including powder (Calcium carbonate/Magnesium carbonate) not more than 50mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Pan / Rose / Guava / Orange / Pineapple / Bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.  <b>For Microbial content product shall comply</b> ISO 11737 / USP 61 &amp; 62.  For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941.  Specification and packaging shall comply BIS/ISO 4074:2015 and shall be square shaped minimum 55mm x 55mm  Conforming ISO 13485:2016/NS-EN ISO 13485:2016</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall pre-qualified be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	Three in Mono-Pack



10	9007	<p><b>Male Latex Condom (1740 Dot):</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pineapple / bubble-gum.  Safety and Standards: Condom shall be tested within international standards i.e., i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941. Product Shall confirm ISO 25841-2017  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	<p><b>Male Latex Condom (1740 Dot): Made from Natural rubber latex</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg including powder (Calcium carbonate/Magnesium carbonate) not more than 50mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Pan / Rose / Guava / Orange / Pineapple / Bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.  <b>For Microbial content product shall comply ISO 11737 / USP 61 &amp; 62.</b>  For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941.  Specification and packaging shall comply BIS/ISO 4074:2015 and shall be square shaped minimum 55mm x 55mm  Conforming ISO 13485:2016/NS-EN ISO 13485:2016</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall pre-qualified be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	Three in Mono-Pack
11	9008	<p><b>Male Latex Condom (Ribbed):</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pineapple / bubble-gum.  Safety and Standards: Condom shall be tested within international standards i.e., i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941. Product Shall confirm ISO 25841-2017  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	<p><b>Male Latex Condom (Ribbed): Made from Natural rubber latex</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg including powder (Calcium carbonate/Magnesium carbonate) not more than 50mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Pan / Rose / Guava / Orange / Pineapple / Bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.  <b>For Microbial content product shall comply ISO 11737 / USP 61 &amp; 62.</b>  For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941.  Specification and packaging shall comply BIS/ISO 4074:2015 and shall be square shaped minimum 55mm x 55mm  Conforming ISO 13485:2016/NS-EN ISO 13485:2016</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall pre-qualified be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	Three in Mono-Pack



12	9009	<p><b>Male Latex Condom (Three in One):</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Paan / Rose / Guava / Orange / Pineapple / bubble-gum.  Safety and Standards: Condom shall be tested within international standards i.e., i.e., ISO 10993-1 (Biological evaluation), ISO 10993-3 (potential genotoxicity, carcinogenicity, or reproductive toxicity), ISO 10993-5 (Cytotoxicity), ISO 10993-10 (Mucosal Irritation &amp; Sensitization), ISO 11737.  For protein test: ISO 12243 &amp; For Nitrosamines: ISO 29941. Product Shall confirm ISO 25841-2017  Specification shall comply BIS/ISO 4074:2014  Conforming QMS 13485:2016/NS-EN ISO 13485:2016  Packaging: Shall comply ISO 25841:2017</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	<p><b>Male Latex Condom (Three in One): Made from Natural rubber latex</b>  Width: 53 ± 2 mm  Length: 180 mm  Thickness: 0.060- 0.080 mm  Lubricant: Silicon Oil; Not less than 350mg including powder (Calcium carbonate/Magnesium carbonate) not more than 50mg  Colour: Pink / Blue / Green / Yellow / Red / Natural / Brown  Flavour: Chocolate / Strawberry / Litchi / Apple Pitch / Fruit Punch / Jasmine / Mango / Banana / Black Grape / Coffee / Hazelnut / Mixed Fruit / Red Bull / Toffee / Bela / Green - apple / Butterscotch / Pan / Rose / Guava / Orange / Pineapple / Bubble-gum.  Safety and Standards: Condom shall be tested within International standards i.e., ISO 10993-5, ISO 10993-10 and ISO 10993-23.  <b>For Microbial content product shall comply ISO 11737 / USP 61 &amp; 62.</b>  For protein test: ISO 12243 / EN 455-3 / ASTM D5712, shall be less than 200 µg / g &amp; For Nitrosamines: ISO 29941.  Specification and packaging shall comply BIS/ISO 4074:2015 and shall be square shaped minimum 55mm x 55mm  Conforming ISO 13485:2016/NS-EN ISO 13485:2016</p> <p>Shelf Life: 3 Years</p> <p>NOTE: Condom shall pre-qualified be free from holes and shall conform Schedule R of Drugs and Cosmetics Act 1940 and Rules there under 1945.  Hazardous Raw Material shall strictly not be Used.</p>	Three in Mono-Pack
13	9012	<p><b>Lumbo-Sacral Support Belt (Waist and back support):</b>  For unisex with flexible back splints, Easy to fit body shape for excellent Immobilization,  Material: Porous elastic webbing, Heat resistant rubber with high modulus of elasticity  Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression, and better sizing flexibility.  ISO 13485 and CE certified</p> <p>Size: Small 28-32 inch  One in mono-pack</p>	<p><b>Lumbo-Sacral Support Belt (Waist and back support):</b>  For unisex with flexible back splints made of plastic/Nylon, Easy to fit the body shape for excellent Immobilization, shall be suitably designed with broad base at bottom to support Lumber 1 bone  Material: Porous buff elastic, Heat resistant rubber with high modulus of elasticity  Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression and better sizing flexibility. Suitably designed to adjust horizontal pressure of 35-40 kg. Velcro Should be of Nylon  ISO 13485 and CE certified</p> <p>Size: Small 28-32 inch</p>	One in Mono-Pack
14	9013	<p><b>Lumbo-Sacral Support Belt (Waist and back support):</b>  For unisex with flexible back splints, Easy to fit body shape for excellent Immobilization,  Material: Porous elastic webbing, Heat resistant rubber with high modulus of elasticity  Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression, and better sizing flexibility.  ISO 13485 and CE certified  Size: Medium 32-36 inch  One in mono-pack</p>	<p><b>Lumbo-Sacral Support Belt (Waist and back support):</b>  For unisex with flexible back splints made of plastic/Nylon, Easy to fit the body shape for excellent Immobilization, shall be suitably designed with broad base at bottom to support Lumber 1 bone  Material: Porous buff elastic, Heat resistant rubber with high modulus of elasticity  Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression and better sizing flexibility. Suitably designed to adjust horizontal pressure of 35-40 kg. Velcro Should be of Nylon  ISO 13485 and CE certified</p> <p>Size: Medium 32-36 inch</p>	One in Mono-Pack
15	9014	<p><b>Lumbo-Sacral Support Belt (Waist and back support):</b>  For unisex with flexible back splints, Easy to fit body shape for excellent Immobilization,  Material: Porous elastic webbing, Heat resistant rubber with high modulus of elasticity</p>	<p><b>Lumbo-Sacral Support Belt (Waist and back support):</b>  For unisex with flexible back splints made of plastic/Nylon, Easy to fit the body shape for excellent Immobilization, shall be suitably designed with broad base at bottom to support Lumber 1 bone</p>	One in Mono-Pack

		Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression, and better sizing flexibility. ISO 13485 and CE certified Size: Large 36-40 inch One in mono-pack	Material: Porous buff elastic, Heat resistant rubber with high modulus of elasticity Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression and better sizing flexibility. Suitably designed to adjust horizontal pressure of 35-40 kg. Velcro Should be of Nylon ISO 13485 and CE certified  Size: Large 36-40 inch	
16	9015	<b>Lumbo-Sacral Support Belt (Waist and back support):</b> For unisex with flexible back splints, Easy to fit body shape for excellent Immobilization, Material: Porous elastic webbing, Heat resistant rubber with high modulus of elasticity Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression, and better sizing flexibility. ISO 13485 and CE certified Size: Extra Large 40-44 inch One in mono-pack	<b>Lumbo-Sacral Support Belt (Waist and back support):</b> For unisex with flexible back splints made of plastic/Nylon, Easy to fit the body shape for excellent Immobilization, shall be suitably designed with broad base at bottom to support Lumber 1 bone Material: Porous buff elastic, Heat resistant rubber with high modulus of elasticity Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression and better sizing flexibility. Suitably designed to adjust horizontal pressure of 35-40 kg. Velcro Should be of Nylon ISO 13485 and CE certified  Size: Extra Large 40-44 inch	One in Mono-Pack
17	9016	<b>Lumbo-Sacral Support Belt (Waist and back support):</b> for unisex with flexible back splints, Easy to fit body shape for excellent Immobilization, Material: Porous elastic webbing, Heat resistant rubber with high modulus of elasticity Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression, and better sizing flexibility. ISO 13485 and CE certified  Size: Double Extra Large 44-48 inch One in mono-pack	<b>Lumbo-Sacral Support Belt (Waist and back support):</b> For unisex with flexible back splints made of plastic/Nylon, Easy to fit the body shape for excellent Immobilization, shall be suitably designed with broad base at bottom to support Lumber 1 bone Material: Porous buff elastic, Heat resistant rubber with high modulus of elasticity Double pull mechanism for secure fitting around the waist, suitably large hook loop panel for controlled compression and better sizing flexibility. Suitably designed to adjust horizontal pressure of 35-40 kg. Velcro Should be of Nylon ISO 13485 and CE certified  Size: Double Extra Large 44-48 inch	One in Mono-Pack
18	9017	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton), Heat resistant rubber with high modulus of elasticity. Hypoallergenic, uniform compression, simple pull-on application.  To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc. One in mono-pack.  Size: Small (18-21 Inch)	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity. Hypoallergenic, uniform compression, simple pull-on application.  To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc.  Size: Small (18-21cms)	Pair in Mono-Pack
19	9018	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton), Heat resistant rubber with high modulus of elasticity. Hypoallergenic, uniform compression, simple pull-on application. To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc. One in mono-pack Size: Medium (21-24 Inch)	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity. Hypoallergenic, uniform compression, simple pull-on application.  To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc.	Pair in Mono-Pack

			Size: Small (21-24cms)	
20	9019	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton), Heat resistant rubber with high modulus of elasticity. hypoallergenic, uniform compression, simple pull-on application. To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc. One in mono-pack Size: Large (24-27 Inch)	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity. Hypoallergenic, uniform compression, simple pull-on application.  To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc. Size: Small (24-27cms)	Pair in Mono-Pack
21	9020	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton), Heat resistant rubber with high modulus of elasticity. hypoallergenic, uniform compression, simple pull-on application. To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc. One in mono-pack Size: Extra Large (27-30 Inch).	<b>Anklet (Tubular Support):</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity. Hypoallergenic, uniform compression, simple pull-on application.  To provide mild compression, warmth & support to the ankle joint. Suitable for pain and inflammation associated with old age arthritis support etc. Size: Small (27-30cms)	Pair in Mono-Pack
22	9021	<b>Elbow Support:</b> Material: Breathable, hypoallergenic, dual-stretch power knit material made of cotton, nylon. Easy to put on and take off; use on left or right elbow Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity. Size: Small 18-25 cm	<b>Elbow Support:</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity, Hypoallergenic, uniform compression, simple pull-on application.  Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity. Size: Small 18-25 cm	Pair in Mono-Pack
23	9022	<b>Elbow Support:</b> Material: Breathable, hypoallergenic, dual-stretch power knit material made of cotton, nylon. Easy to put on and take off; use on left or right elbow Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity.  Size: Medium 25-30 cm	<b>Elbow Support:</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity, Hypoallergenic, uniform compression, simple pull-on application.  Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity. Size: Medium 25-30 cm	Pair in Mono-Pack
24	9023	<b>Elbow Support:</b> Material: Breathable, hypoallergenic, dual-stretch power knit material made of cotton, nylon. Easy to put on and take off; use on left or right elbow, Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity. Size: Large 30-35 cm	<b>Elbow Support:</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity, Hypoallergenic, uniform compression, simple pull-on application.  Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity. Size: Large 30-35 cm	Pair in Mono-Pack

25	9024	<b>Elbow Support:</b> Material: Breathable, hypoallergenic, dual-stretch power knit material made of cotton, nylon. Easy to put on and take off; use on left or right elbow Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity. Size: Extra Large 35-40 cm	<b>Elbow Support:</b> Four way stretchable, bi-layered fabric (outer layer made of nylon and inner layer made of fine grade cotton) in ratio of 60% Nylon : 40% Cotton, Heat resistant rubber with high modulus of elasticity, Hypoallergenic, uniform compression, simple pull-on application.  Helps provide compression, relief, and protection. Dimensionally fit to provide uniform compression. Heat resistant rubber with high modulus of elasticity.  Size: Extra Large 35-40 cm	Pair in Mono-Pack
26	9025	<b>Combined Dressing Pad</b> Each sterile dressing pad consist of highly absorbent layer of fleece enclosed in a soft and comfortable non-woven fabric to prevent pressure areas forming. Absorbent Pad made from cotton surrounded by absorbent gauze, sterile, free from any brightener. Properly Wrapped. Size: 10x10 cm	<b>Combined Dressing Pad</b> Each sterile dressing pad consist of highly absorbent layer of fleece enclosed in a soft and comfortable non-woven fabric to prevent pressure areas forming. Absorbent Pad made from cotton surrounded by non-woven fabric, sterile, free from any brightener. Each pad shall be properly Wrapped. Size: 10x10 cm <b>Weight: Minimum 5 g per pad</b>	Five in mono-pack
27	9026	<b>Combined Dressing Pad</b> Each sterile dressing pad consist of highly absorbent layer of fleece enclosed in a soft and comfortable non-woven fabric to prevent pressure areas forming. Absorbent Pad made from cotton surrounded by absorbent gauze, sterile, free from any brightener. Properly Wrapped. Five in mono-pack Size: 10x20cm	<b>Combined Dressing Pad</b> Each sterile dressing pad consist of highly absorbent layer of fleece enclosed in a soft and comfortable non-woven fabric to prevent pressure areas forming. Absorbent Pad made from cotton surrounded by non-woven fabric, sterile, free from any brightener. Each pad shall be properly Wrapped. Five in mono-pack Size: 10x20cm <b>Weight: Minimum 10 g per pad</b>	Five in mono-pack
28	9028	<b>Medical Steam Vaporizer:</b> Model: Sleek style Tank Capacity: 400-500ml Comfortable to use in Cough Sinus Nose Console, Steamer for Facial, Cold and Cough, Steam Machine for Face (vaporizer) Steam Vaporizer, Nose Steamer, Cough Steamer, Nozzle Inhaler & Nose vaporizer machine for cold and cough Body / Material: 100% high grade plastic body with double wall protection to prevent shocks and heat injury. Sufficient space for easy cleaning. Electric Cable: ISI certified 1.5 meter long, Plug shall be so designed to save from any shock while plugging in. Water boiling time: 3-5 minutes Warranty: One year Accessory: Vaporizer with nasal aspirator and steaming mask. Product shall be CE/ISO/FDA approved.	<b>Medical Steam Vaporizer:</b> Model: Sleek style / Standard model Tank Capacity: 400-500ml Comfortable to use in Cough Sinus Nose Console, Steamer for Facial, Cold and Cough, Steam Machine for Face (vaporizer) Steam Vaporizer, Nose Steamer, Cough Steamer, Nozzle Inhaler & Nose vaporizer machine for cold and cough Body / Material: 100% high grade plastic body with double wall protection to prevent shocks and heat injury. Sufficient space for easy cleaning. Electric Cable: ISI certified 1.5 meter long, Plug shall be so designed to save from any shock while plugging in. Water boiling time: 3-5 minutes Warranty: One year Accessory: Vaporizer with Nasal Aspirator and Steaming mask / Facial Attachment. Product shall be CE/ISO/FDA approved.	One in Mono-Pack

### **Part- C**

The following packaging type is hereby authorized against item code:

Sl. No.	Reference Clause/Annexure No.	Item code and Generic Specification	Packaging specification
1	Clause 21, ANNEXURE - VIII and ANNEXURE – XII	8133; Digital Blood Pressure Instrument 8136; Electric Nebulizer Machine 9028; Medical Steam Vaporizer	<ul style="list-style-type: none"> <li>➤ Single unit (device) shall be packed in 3ply corrugated FBB/SBS box (mono-carton not less than 350 GSM) to protect the product from any damage during transit.</li> <li>➤ Product shall be packed in foam-based packing with density not less than 18kg/m<sup>3</sup>.</li> <li>➤ Thickness must not be less than 1 mm for item code 8133 &amp; 9028. Item code 8136 shall be packed with foam-based cover of not less than 5 mm thickness.</li> </ul>
2		9003; Female Natural Latex Condom. 9004,9005, 9006, 9007, 9008, 9009; Male Latex Condom	<ul style="list-style-type: none"> <li>➤ Packaging shall comply ISO 25841:2017 for Female condom and ISO 4074:2015 for Male condom.</li> <li>➤ Mono-pack shall not be of less than 350 GSM and shall be packed in compliance to ISO 25841:2017 and ISO 4074:2015 respectively.</li> </ul>

Note: bidders must consider all other conditions mentioned in the tender document besides the above modifications.

All other contents of tender document remain unaltered. Bidders are requested to quote their rates considering all the terms and condition of the tender document including Amendment no. 1 dated 26.07.2022 and Amendment no. 2 dated 05.08.2022.

Sd/-  
DGM (Procurement & Quality)  
For & on behalf of PMBI  
Ph: 011-49431811  
Email: [dgm.pnqc@janaushadhi.gov.in](mailto:dgm.pnqc@janaushadhi.gov.in)